

Serial No. C 95366

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



Members of the
UK Weighing Federation
Membership No. 1070

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

MIDLAND METALS WORKS LTD
5 Bishop Meadow Rd
Leigh
LE11 5RE
LOCATION
platform scale

Machine Details

Cust ref	120610
Make	Bilanciai
Model	E302
Ser No	PL97063
Capacity	3000 kg
e/d	11 kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
212/13	9445	05/04/22

Cert Interval	12 mths
Next Cert	04 2022

Tolerance +/-	kg
0-500	1
500-2000	2
2000-3000	3

Comments ☒
.....
.....
.....
.....

2	4
1	3

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
20	20	0	20	0
1000	1000	0	1000	0
2000	1999	-1	1999	-1
3000	2999	-1	2999	-1

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
2000	1999	2000	2000

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 400

Reading	Dev	Reading	Dev
1 400	0	2 399	-1
3 400	0	4 399	-1
5		6	
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature

Name

Date 29/04/2022

Serial No. C 93905

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

midland metals Recycling
5 Bishop Road
Loughborough
LE11 5RE
LOCATION

Machine Details

Cust ref weighbridge
Make Euro Decu
Model D410
Ser No 129 759
Capacity 50 1000 / kg
e/d 120 / kg
Working cap

Test Weights

Set No.	Cert No.	Date Cal
1-20	TR11408	16.12.21

Cert Interval	yearly
Next Cert	Feb 2023

Tolerance +/-	kg
0-10000	70
10000-40000	40
40000-50000	60

Comments

2	4	6
1	3	5

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
2000	2000	0	0	0
8000	8000	0	2000	0
12000	12000	0	8000	0
14000	14000	0	12000	0
18000	18000	0	14000	0
19600	19600	0	19600	0
29600	29600	0	29600	0
37600	37600	0	37600	0
42360	42360	0	42360	0

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
42360	42360	42360	—

Repeatability

Load	Ind	Ind	Ind

Eccentric Load

Reading	Dev	Reading	Dev
1 6800		2 6800	
3 6800		4 6800	
5 6800		6 6800	
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature S Vardy

Name S. Vardy

Date 25/2/22